
Midmark M11 - v1.0.5
Total Cycles: 1617

Name: RITTER2
Sterilizer ID: V2313228
10/18/2022 12:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.1 F 0.0
2:00 174.1 F 3.0
4:00 200.7 F 8.0
6:00 224.2 F 10.4
8:00 227.8 F 9.7
10:00 234.6 F 8.7
12:00 242.6 F 12.2
14:00 244.8 F 13.4
16:00 258.3 F 20.5
18:00 270.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.1 F 29.8
2:00 271.4 F 28.9
3:00 272.4 F 29.8
4:00 271.8 F 29.2

Min 271.1 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 241.1 F 0.0
10:00 228.0 F 0.0
15:00 233.3 F 0.0
20:00 220.7 F 0.0
25:00 234.8 F 0.0
Dry Cycle Aborted

FILLING: 3:09
HEATING: 18:42
STERILIZING: 4:00
VENTING: 3:14
DRYING: 27:08
TOTAL CYCLE: 00:56:12

10/18/2022 01:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____