
Midmark M11 - v1.0.5
Total Cycles: 2038

Name: RITTER2
Sterilizer ID: V2313228
02/13/2023 08:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 76.3 F 0.0
4:00 80.8 F 0.0
6:00 108.6 F 0.0
8:00 148.9 F 0.0
10:00 174.9 F 1.6
12:00 192.8 F 4.2
14:00 221.0 F 7.1
16:00 224.5 F 10.9
18:00 229.3 F 7.9
20:00 229.6 F 10.6
22:00 234.6 F 8.3
24:00 242.6 F 13.1
26:00 241.9 F 11.6
28:00 254.0 F 17.7
30:00 264.6 F 24.2
32:00 271.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.8 F 29.6
1:00 271.8 F 29.7
2:00 271.9 F 30.1
3:00 272.4 F 30.0
4:00 272.6 F 29.9

Min 271.6 F 29.5
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 219.6 F 0.0
10:00 205.0 F 0.0
15:00 197.9 F 0.0
20:00 195.5 F 0.0
25:00 194.0 F 0.0
30:00 193.7 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:19
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:10:23
CYCLE COMPLETE

02/13/2023 09:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____