

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1450

Name: RITTER 1  
Sterilizer ID: V2178198  
12/13/2021 12:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.6 F 0.0  
2:00 154.7 F 1.7  
4:00 176.7 F 4.7  
6:00 193.3 F 8.8  
8:00 215.3 F 10.8  
10:00 224.3 F 8.2  
12:00 237.1 F 13.3  
14:00 237.3 F 10.4  
16:00 235.2 F 8.1  
18:00 249.3 F 15.1  
20:00 261.0 F 21.7  
22:00 271.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 271.7 F 29.0  
2:00 272.7 F 29.3  
3:00 271.7 F 28.6  
4:00 271.5 F 29.0

Min 271.2 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 218.5 F 0.0  
10:00 239.8 F 0.0  
15:00 236.4 F 0.0  
20:00 226.8 F 0.0  
25:00 226.3 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 22:29  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:02  
CYCLE COMPLETE

12/13/2021 01:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_