
Midmark M11 - v1.0.0
Total Cycles: 3947

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2023 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.0 F 0.0
2:00 155.7 F 3.3
4:00 179.4 F 8.0
6:00 221.4 F 11.6
8:00 225.1 F 7.6
10:00 238.3 F 13.3
12:00 238.5 F 10.2
14:00 235.1 F 8.1
16:00 251.0 F 15.5
18:00 263.7 F 22.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.5
2:00 271.9 F 28.4
3:00 272.6 F 28.5
4:00 271.6 F 28.1

Min 271.5 F 27.9
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 245.9 F 0.0
10:00 230.0 F 0.0
15:00 231.7 F 0.0
20:00 227.3 F 0.0
25:00 229.2 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

10/25/2023 02:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____