
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001995
14:51 09/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:46
STERILIZING 3:52
136.0 C 220kPa 3:52
135.9 C 221kPa 4:22
136.4 C 226kPa 4:52
136.4 C 225kPa 5:22
136.2 C 223kPa 5:52
136.4 C 225kPa 6:22
136.3 C 224kPa 6:52
136.4 C 225kPa 7:22
136.4 C 225kPa 7:22
VENTING 7:22
AIR DRYING 8:16
CYCLE COMPLETE 8:23

Digital Signature #
1748E554EEBFD3C5
