

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4024

Name: RITTER 1  
Sterilizer ID: V2178198  
11/15/2023 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.3 F 0.0  
2:00 162.4 F 2.3  
4:00 186.3 F 6.0  
6:00 195.4 F 10.2  
8:00 216.6 F 4.0  
10:00 224.3 F 8.4  
12:00 234.0 F 13.1  
14:00 235.4 F 9.6  
16:00 240.0 F 9.5  
18:00 246.5 F 13.6  
20:00 258.2 F 19.9  
22:00 268.9 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.7 F 29.1  
2:00 271.7 F 28.5  
3:00 271.9 F 29.0  
4:00 272.5 F 29.0

Min 271.3 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 217.2 F 0.0  
10:00 221.3 F 0.0  
15:00 224.1 F 0.0  
20:00 225.3 F 0.0  
25:00 226.5 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 23:10  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:01:51  
CYCLE COMPLETE

11/15/2023 03:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_