

-----  
Midmark M11 - v1.0.5  
Total Cycles: 991

Name: RITTER2  
Sterilizer ID: V2313228  
04/14/2022 02:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.5 F 0.0  
2:00 176.4 F 0.3  
4:00 186.1 F 2.2  
6:00 208.9 F 7.2  
8:00 223.8 F 13.0  
10:00 232.6 F 9.7  
12:00 235.8 F 9.0  
14:00 241.2 F 11.1  
16:00 247.7 F 14.6  
18:00 260.8 F 22.0  
20:00 272.6 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.4 F 29.5  
2:00 272.1 F 29.8  
3:00 272.1 F 29.2  
4:00 272.4 F 29.9

Min 271.6 F 29.0  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
5:00 208.1 F 0.0  
10:00 215.8 F 0.0  
15:00 217.2 F 0.0  
20:00 222.9 F 0.0  
25:00 223.2 F 0.0  
30:00 227.2 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:41  
CYCLE COMPLETE

04/14/2022 03:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_