
Midmark M11 - v1.0.5
Total Cycles: 1171

Name: RITTER2
Sterilizer ID: V2313228
06/08/2022 07:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.2 F 0.0
2:00 77.0 F 0.0
4:00 80.7 F 0.0
6:00 93.7 F 0.0
8:00 128.6 F 1.5
10:00 158.3 F 4.8
12:00 191.2 F 8.5
14:00 214.9 F 12.5
16:00 218.9 F 6.6
18:00 230.1 F 10.7
20:00 229.7 F 8.2
22:00 239.8 F 12.4
24:00 238.9 F 10.8
26:00 251.2 F 16.9
28:00 262.0 F 23.2
30:00 272.3 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 271.5 F 29.7
2:00 272.5 F 30.2
3:00 272.0 F 29.6
4:00 271.2 F 29.3

Min 271.2 F 29.2
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 193.2 F 0.0
10:00 200.9 F 0.0
15:00 199.4 F 0.0
20:00 200.1 F 0.0
25:00 200.2 F 0.0
30:00 198.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:22
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:18
CYCLE COMPLETE

06/08/2022 08:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____