
Midmark M11 - v1.0.5
Total Cycles: 572

Name: RITTER2
Sterilizer ID: V2313228
11/24/2021 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.4 F 0.0
2:00 175.3 F 1.6
4:00 176.0 F 5.4
6:00 195.6 F 10.5
8:00 213.7 F 5.2
10:00 227.9 F 10.6
12:00 231.5 F 7.7
14:00 241.0 F 11.2
16:00 245.5 F 13.7
18:00 257.9 F 20.4
20:00 269.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.6 F 29.9
2:00 271.5 F 29.1
3:00 272.1 F 30.1
4:00 272.3 F 29.7

Min 271.3 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 225.0 F 0.0
10:00 233.5 F 0.0
15:00 226.0 F 0.0
20:00 223.7 F 0.0
25:00 223.9 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 21:03
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

11/24/2021 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____