

-----  
Midmark M11 - v1.0.5  
Total Cycles: 305

Name: RITTER2  
Sterilizer ID: V2313228  
08/24/2021 07:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.0 F 0.0  
2:00 155.2 F 1.8  
4:00 169.6 F 5.5  
6:00 189.4 F 10.2  
8:00 217.5 F 4.7  
10:00 231.3 F 9.9  
12:00 233.3 F 8.4  
14:00 244.3 F 13.2  
16:00 243.1 F 12.2  
18:00 256.0 F 19.0  
20:00 267.4 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.2 F 29.2  
2:00 271.9 F 29.5  
3:00 272.0 F 28.9  
4:00 272.4 F 29.6

Min 271.3 F 28.7  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
5:00 222.9 F 0.0  
10:00 206.1 F 0.0  
15:00 220.2 F 0.0  
20:00 213.3 F 0.0  
25:00 214.2 F 0.0  
30:00 209.9 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 21:21  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:00:14  
CYCLE COMPLETE

08/24/2021 09:00 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02