
Midmark M11 - v1.0.5
Total Cycles: 3602

Name: RITTER2
Sterilizer ID: V2313228
05/09/2024 10:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 144.8 F 1.2
4:00 147.4 F 3.8
6:00 170.0 F 7.2
8:00 192.9 F 11.3
10:00 214.6 F 4.8
12:00 227.2 F 9.4
14:00 232.5 F 10.6
16:00 237.9 F 11.1
18:00 235.7 F 9.2
20:00 248.9 F 15.3
22:00 259.8 F 21.5
24:00 269.8 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.8
1:00 272.1 F 29.7
2:00 272.7 F 29.9
3:00 272.6 F 30.2
4:00 271.9 F 30.3

Min 271.6 F 29.7
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 201.9 F 0.0
10:00 192.1 F 0.0
15:00 190.9 F 0.0
20:00 190.1 F 0.0
25:00 189.1 F 0.0
30:00 200.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 25:42
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:04:23
CYCLE COMPLETE

05/09/2024 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____