
Midmark M11 - v1.0.5
Total Cycles: 1447

Name: RITTER2
Sterilizer ID: V2313228
09/02/2022 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.9 F 0.0
2:00 195.6 F 2.8
4:00 196.5 F 7.2
6:00 218.5 F 12.4
8:00 226.4 F 8.0
10:00 236.5 F 13.5
12:00 238.1 F 11.1
14:00 237.5 F 9.6
16:00 250.9 F 16.2
18:00 262.3 F 22.8
20:00 272.6 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.4 F 29.6
2:00 271.4 F 29.0
3:00 272.4 F 29.7
4:00 271.7 F 29.0

Min 271.3 F 28.9
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 222.2 F 0.0
10:00 233.2 F 0.0
15:00 230.6 F 0.0
20:00 228.8 F 0.0
25:00 226.0 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

09/02/2022 10:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____