
Midmark M11 - v1.0.0
Total Cycles: 1163

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2021 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.4 F 0.0
2:00 192.6 F 3.4
4:00 211.8 F 8.6
6:00 222.5 F 8.2
8:00 233.7 F 9.9
10:00 237.0 F 8.4
12:00 244.8 F 12.8
14:00 244.4 F 12.5
16:00 257.1 F 19.2
18:00 268.1 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.5 F 29.2
2:00 271.8 F 28.5
3:00 272.3 F 29.2
4:00 272.3 F 28.8

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 229.3 F 0.0
10:00 229.2 F 0.0
15:00 225.2 F 0.0
20:00 228.7 F 0.0
25:00 225.8 F 0.0
Dry Cycle Aborted

FILLING: 2:07
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:11
DRYING: 27:39
TOTAL CYCLE: 00:56:06

09/13/2021 12:02 PM

ACCEPTED: X

REJECTED:

Operator: U01