
Midmark M11 - v1.0.5
Total Cycles: 843

Name: RITTER2
Sterilizer ID: V2313228
02/28/2022 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.6 F 0.0
2:00 176.0 F 2.9
4:00 195.1 F 7.8
6:00 215.9 F 12.3
8:00 227.5 F 9.5
10:00 233.4 F 8.4
12:00 242.7 F 12.0
14:00 246.4 F 13.7
16:00 259.5 F 21.0
18:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.4
2:00 271.5 F 28.7
3:00 272.3 F 29.6
4:00 272.2 F 29.2

Min 271.3 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 215.0 F 0.0
10:00 227.5 F 0.0
15:00 205.5 F 0.0
20:00 203.3 F 0.0
25:00 207.5 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 18:42
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

02/28/2022 12:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____