
Midmark M11 - v1.0.5
Total Cycles: 2816

Name: RITTER2
Sterilizer ID: V2313228
08/30/2023 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.4 F 0.0
2:00 172.4 F 0.0
4:00 194.7 F 2.7
6:00 214.6 F 7.5
8:00 230.6 F 12.8
10:00 234.4 F 9.1
12:00 236.8 F 9.7
14:00 245.1 F 13.4
16:00 244.3 F 12.7
18:00 257.3 F 19.7
20:00 268.8 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.4 F 29.3
2:00 271.3 F 28.9
3:00 272.4 F 29.4
4:00 271.4 F 28.6

Min 271.1 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 235.9 F 0.0
10:00 221.7 F 0.0
15:00 235.2 F 0.0
20:00 205.2 F 0.0
25:00 190.1 F 0.0
30:00 183.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

08/30/2023 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____