

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000174  
13:39 16/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:35  
PRESSURIZING 4:36  
STERILIZING 6:06  
136.4 C 225kPa 6:06  
136.4 C 226kPa 6:37  
136.4 C 225kPa 7:07  
136.1 C 223kPa 7:37  
136.3 C 225kPa 8:07  
136.3 C 225kPa 8:37  
136.4 C 225kPa 9:07  
136.4 C 225kPa 9:37  
136.4 C 225kPa 9:37  
VENTING 9:37  
AIR DRYING 10:39  
CYCLE COMPLETE 25:39  
-----

Digital Signature #  
F53B2DE906E6C93D  
-----