

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002523  
10:14 20/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:55  
PRESSURIZING 2:00  
STERILIZING 3:08  
136.4 C 224kPa 3:08  
136.4 C 226kPa 3:38  
136.0 C 222kPa 4:08  
136.2 C 224kPa 4:38  
136.4 C 226kPa 5:08  
136.5 C 224kPa 5:38  
135.9 C 222kPa 6:08  
136.4 C 224kPa 6:38  
136.4 C 224kPa 6:38  
VENTING 6:38  
AIR DRYING 7:35  
CYCLE COMPLETE 16:00  
-----

Digital Signature #  
85CE31CD1A117CF2  
-----