
Midmark M11 - v1.0.5
Total Cycles: 1780

Name: RITTER2
Sterilizer ID: V2313228
11/28/2022 03:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.7 F 0.0
2:00 207.4 F 4.0
4:00 210.4 F 9.9
6:00 225.3 F 7.1
8:00 237.7 F 12.7
10:00 239.5 F 11.4
12:00 240.2 F 10.9
14:00 254.2 F 18.2
16:00 266.5 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.4 F 29.6
2:00 272.2 F 29.9
3:00 272.1 F 29.3
4:00 272.4 F 30.0

Min 271.6 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 222.5 F 0.0
10:00 228.9 F 0.0
15:00 229.8 F 0.0
20:00 227.6 F 0.0
25:00 226.4 F 0.0
30:00 224.4 F 0.0

Drying Complete

FILLING: 3:05
HEATING: 17:28
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:46
CYCLE COMPLETE

11/28/2022 04:36 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____