
Midmark M11 - v1.0.5
Total Cycles: 2618

Name: RITTER2
Sterilizer ID: V2313228
07/06/2023 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.6 F 0.0
2:00 195.7 F 3.6
4:00 201.8 F 8.8
6:00 230.4 F 9.2
8:00 231.1 F 10.8
10:00 235.1 F 8.5
12:00 240.5 F 10.2
14:00 247.7 F 14.4
16:00 260.3 F 21.4
18:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.5
1:00 271.9 F 29.6
2:00 271.9 F 29.5
3:00 272.0 F 29.4
4:00 272.2 F 29.6

Min 271.8 F 29.4
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 223.8 F 0.0
10:00 237.3 F 0.0
15:00 229.0 F 0.0
20:00 228.3 F 0.0
25:00 229.6 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

07/06/2023 01:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____