
Midmark M11 - v1.0.0
Total Cycles: 1047

Name: RITTER 1
Sterilizer ID: V2178198
07/22/2021 12:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.4 F 0.0
2:00 197.1 F 3.9
4:00 221.3 F 9.5
6:00 226.3 F 6.7
8:00 238.4 F 11.2
10:00 236.2 F 8.8
12:00 239.5 F 9.6
14:00 249.7 F 14.8
16:00 262.5 F 22.1
18:00 271.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.2 F 28.3
2:00 272.4 F 28.7
3:00 271.7 F 28.0
4:00 272.4 F 28.7

Min 271.3 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 232.5 F 0.0
10:00 230.7 F 0.0
15:00 199.3 F 0.0
20:00 185.0 F 0.0
25:00 184.4 F 0.0
30:00 198.1 F 0.0
Drying Complete

FILLING: 3:42
HEATING: 18:06
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

07/22/2021 01:45 PM
ACCEPTED: X
REJECTED:
Operator: U02