
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001016
16:08 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:01
STERILIZING 3:01
136.5 C 233kPa 3:01
136.3 C 233kPa 3:31
136.3 C 232kPa 4:01
136.4 C 230kPa 4:31
136.3 C 233kPa 5:01
136.2 C 232kPa 5:31
136.1 C 231kPa 6:01
136.3 C 233kPa 6:31
136.3 C 233kPa 6:31
VENTING 6:31
AIR DRYING 7:22
CYCLE COMPLETE 7:25

Digital Signature #
ED281B241158264E
