
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008377
14:19 05/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:12
STERILIZING 4:15
136.4 C 221kPa 4:15
136.5 C 225kPa 4:45
136.1 C 219kPa 5:15
136.1 C 222kPa 5:45
136.2 C 222kPa 6:15
136.3 C 224kPa 6:45
136.4 C 223kPa 7:15
136.3 C 222kPa 7:45
136.3 C 222kPa 7:45
VENTING 7:45
AIR DRYING 8:41
CYCLE COMPLETE 38:41

Digital Signature #
95AAE5F2B36AD77D
