
Midmark M11 - v1.0.5
Total Cycles: 2846

Name: RITTER2
Sterilizer ID: V2313228
09/07/2023 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.7 F 0.0
2:00 131.0 F 0.0
4:00 145.3 F 1.4
6:00 179.2 F 4.7
8:00 191.3 F 8.5
10:00 216.0 F 12.7
12:00 223.2 F 7.5
14:00 232.6 F 12.1
16:00 233.3 F 9.1
18:00 238.9 F 10.5
20:00 245.4 F 13.7
22:00 256.6 F 19.8
24:00 267.3 F 26.4
26:00 272.6 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.6
1:00 271.7 F 29.5
2:00 272.0 F 29.5
3:00 271.9 F 29.6
4:00 272.0 F 29.8

Min 271.5 F 29.4
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 205.2 F 0.0
10:00 208.2 F 0.0
15:00 201.2 F 0.0
20:00 197.5 F 0.0
25:00 198.3 F 0.0
30:00 202.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:10
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:04:46
CYCLE COMPLETE

09/07/2023 03:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____