
Midmark M11 - v1.0.0
Total Cycles: 3334

Name: RITTER 1
Sterilizer ID: V2178198
06/05/2023 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.4 F 0.0
2:00 197.5 F 1.0
4:00 218.1 F 6.6
6:00 229.5 F 12.2
8:00 231.6 F 8.3
10:00 241.4 F 11.9
12:00 243.2 F 12.2
14:00 241.8 F 11.1
16:00 255.4 F 18.0
18:00 267.1 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.8
2:00 271.5 F 28.0
3:00 272.2 F 28.8
4:00 272.3 F 28.6

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 220.8 F 0.0
10:00 231.6 F 0.0
15:00 230.8 F 0.0
20:00 224.5 F 0.0
25:00 223.5 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

06/05/2023 11:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____