
Midmark M11 - v1.0.5
Total Cycles: 1214

Name: RITTER2
Sterilizer ID: V2313228
06/22/2022 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.3 F 0.0
2:00 178.1 F 3.7
4:00 184.5 F 8.0
6:00 208.3 F 13.0
8:00 222.1 F 7.9
10:00 235.3 F 13.2
12:00 237.8 F 10.9
14:00 237.2 F 9.6
16:00 251.1 F 16.5
18:00 262.7 F 23.3
20:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 29.9
2:00 271.5 F 29.1
3:00 272.5 F 30.1
4:00 272.2 F 29.6

Min 271.3 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 216.2 F 0.0
10:00 237.6 F 0.0
15:00 229.2 F 0.0
20:00 225.9 F 0.0
25:00 224.7 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

06/22/2022 02:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____