
Midmark M11 - v1.0.5
Total Cycles: 1931

Name: RITTER2
Sterilizer ID: V2313228
01/12/2023 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.1 F 0.0
2:00 203.5 F 4.2
4:00 216.8 F 9.9
6:00 222.0 F 6.5
8:00 231.4 F 11.8
10:00 237.5 F 10.0
12:00 235.5 F 9.0
14:00 251.5 F 16.5
16:00 263.8 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.4 F 30.1
2:00 272.1 F 29.4
3:00 272.9 F 30.0
4:00 272.1 F 29.6

Min 271.0 F 28.7
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 236.2 F 0.0
10:00 233.5 F 0.0
15:00 225.9 F 0.0
20:00 219.6 F 0.0
25:00 219.1 F 0.0
30:00 217.2 F 0.0

Drying Complete

FILLING: 2:16
HEATING: 17:52
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:17
CYCLE COMPLETE

01/12/2023 04:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____