
Midmark M11 - v1.0.5
Total Cycles: 2866

Name: RITTER2
Sterilizer ID: V2313228
09/12/2023 07:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 72.3 F 0.0
4:00 88.4 F 0.0
6:00 103.8 F 0.0
8:00 142.1 F 1.2
10:00 163.8 F 3.9
12:00 187.8 F 7.0
14:00 205.2 F 10.6
16:00 215.3 F 6.8
18:00 223.2 F 8.7
20:00 233.8 F 13.0
22:00 234.1 F 9.3
24:00 239.6 F 10.7
26:00 244.4 F 12.9
28:00 256.2 F 19.0
30:00 266.4 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 272.4 F 29.5
2:00 271.1 F 29.1
3:00 272.3 F 29.5
4:00 271.0 F 29.0

Min 271.0 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 197.6 F 0.0
10:00 202.6 F 0.0
15:00 204.2 F 0.0
20:00 213.3 F 0.0
25:00 216.0 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:40
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:09:49
CYCLE COMPLETE

09/12/2023 09:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____