
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001435
16:16 14/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:15
STERILIZING 4:27
136.4 C 226kPa 4:27
136.4 C 225kPa 4:57
136.3 C 224kPa 5:27
136.3 C 224kPa 5:57
136.3 C 224kPa 6:27
136.3 C 225kPa 6:57
136.2 C 224kPa 7:27
136.1 C 225kPa 7:57
136.1 C 225kPa 7:57
VENTING 7:57
AIR DRYING 8:48
CYCLE COMPLETE 9:45

Digital Signature #
F81785B8D3067CB4
