
Midmark M11 - v1.0.5
Total Cycles: 1301

Name: RITTER2
Sterilizer ID: V2313228
07/14/2022 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.0 F 0.0
2:00 130.0 F 0.0
4:00 155.7 F 0.8
6:00 182.4 F 3.8
8:00 206.3 F 7.7
10:00 225.9 F 12.5
12:00 228.3 F 7.5
14:00 238.5 F 12.4
16:00 237.6 F 9.9
18:00 237.4 F 9.2
20:00 249.0 F 14.9
22:00 260.5 F 21.4
24:00 271.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 29.0
2:00 272.6 F 29.5
3:00 271.4 F 28.6
4:00 272.5 F 29.3

Min 271.3 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.7 F 0.0
5:00 219.7 F 0.0
10:00 217.6 F 0.0
15:00 222.5 F 0.0
20:00 231.5 F 0.0
25:00 224.9 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 24:41
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:03:27
CYCLE COMPLETE

07/14/2022 04:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____