
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000329
16:05 17/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:40
STERILIZING 4:45
136.4 C 224kPa 4:45
136.4 C 225kPa 5:15
136.3 C 221kPa 5:45
136.4 C 225kPa 6:15
136.4 C 224kPa 6:45
136.4 C 223kPa 7:15
136.1 C 222kPa 7:45
136.2 C 220kPa 8:15
136.2 C 220kPa 8:15
VENTING 8:15
AIR DRYING 9:12
CYCLE COMPLETE 9:27

Digital Signature #
F6B382D7C85B6CBA
