

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4016

Name: RITTER2  
Sterilizer ID: V2313228  
10/10/2024 08:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 71.5 F 0.0  
4:00 78.3 F 0.0  
6:00 115.6 F 0.1  
8:00 157.1 F 0.2  
10:00 187.6 F 0.5  
12:00 205.0 F 3.6  
14:00 214.4 F 6.7  
16:00 224.5 F 10.2  
18:00 230.1 F 9.3  
20:00 236.1 F 10.5  
22:00 235.4 F 8.5  
24:00 244.0 F 12.1  
26:00 245.5 F 13.1  
28:00 257.5 F 19.6  
30:00 268.4 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.5 F 29.0  
2:00 272.1 F 29.2  
3:00 272.1 F 28.7  
4:00 272.6 F 29.4  
  
Min 271.5 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
5:00 208.4 F 0.0  
10:00 203.6 F 0.0  
15:00 201.0 F 0.0  
20:00 202.4 F 0.0  
25:00 205.0 F 0.0  
30:00 208.0 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:13  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:09:24  
CYCLE COMPLETE

10/10/2024 09:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_