
Midmark M11 - v1.0.0
Total Cycles: 4863

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2024 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.3 F 0.0
2:00 99.0 F 0.0
4:00 116.0 F 0.2
6:00 145.2 F 0.4
8:00 173.8 F 0.6
10:00 197.2 F 1.5
12:00 210.7 F 5.0
14:00 220.5 F 8.2
16:00 230.4 F 11.8
18:00 229.1 F 7.1
20:00 241.2 F 12.2
22:00 238.4 F 9.9
24:00 237.8 F 8.9
26:00 249.3 F 14.6
28:00 260.5 F 20.8
30:00 270.5 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.5 F 28.7
2:00 271.8 F 28.1
3:00 271.5 F 28.1
4:00 272.4 F 28.6

Min 271.3 F 27.8
Max 272.5 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 207.7 F 0.0
10:00 206.2 F 0.0
15:00 212.0 F 0.0
20:00 223.6 F 0.0
25:00 233.2 F 0.0
30:00 232.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:54
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:09:21
CYCLE COMPLETE

08/29/2024 02:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____