
Midmark M11 - v1.0.5
Total Cycles: 3450

Name: RITTER2
Sterilizer ID: V2313228
03/12/2024 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 75.9 F 0.0
4:00 82.0 F 0.1
6:00 106.0 F 1.2
8:00 144.3 F 3.8
10:00 170.5 F 6.8
12:00 185.1 F 10.6
14:00 209.4 F 5.1
16:00 216.2 F 6.7
18:00 226.9 F 10.8
20:00 230.4 F 7.5
22:00 240.8 F 12.2
24:00 239.2 F 10.5
26:00 251.8 F 16.7
28:00 262.7 F 23.0
30:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.5
2:00 271.5 F 29.1
3:00 272.5 F 29.5
4:00 271.5 F 29.0

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.5 F 0.0
5:00 230.6 F 0.0
10:00 209.9 F 0.0
15:00 207.9 F 0.0
20:00 205.8 F 0.0
25:00 207.0 F 0.0
30:00 208.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:23
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:08:35
CYCLE COMPLETE

03/12/2024 10:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____