

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000570  
14:37 06/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:18  
PRESSURIZING 2:28  
STERILIZING 3:38  
136.0 C 230kPa 3:38  
136.4 C 233kPa 4:08  
136.2 C 232kPa 4:38  
136.4 C 232kPa 5:08  
136.3 C 230kPa 5:38  
136.3 C 232kPa 6:08  
136.3 C 233kPa 6:38  
136.4 C 233kPa 7:08  
136.4 C 233kPa 7:08  
VENTING 7:08  
AIR DRYING 8:03  
CYCLE COMPLETE 23:03

-----  
Digital Signature #  
3FFAEDFC747F5C85  
-----