

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2075

Name: RITTER 1  
Sterilizer ID: V2178198  
06/29/2022 01:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
2:00 186.1 F 3.8  
4:00 200.5 F 8.4  
6:00 216.8 F 13.5  
8:00 224.3 F 7.6  
10:00 237.2 F 12.9  
12:00 237.7 F 10.0  
14:00 235.7 F 8.3  
16:00 249.5 F 15.0  
18:00 261.5 F 21.7  
20:00 272.4 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.0  
2:00 271.7 F 28.3  
3:00 271.9 F 28.9  
4:00 272.5 F 28.9  
  
Min 271.3 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
5:00 213.9 F 0.0  
10:00 222.7 F 0.0  
15:00 232.6 F 0.0  
20:00 230.0 F 0.0  
25:00 224.8 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:30  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:58  
CYCLE COMPLETE

06/29/2022 02:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_