
Midmark M11 - v1.0.5
Total Cycles: 2500

Name: RITTER2
Sterilizer ID: V2313228
06/12/2023 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.8 F 0.0
2:00 174.6 F 3.9
4:00 184.4 F 8.1
6:00 204.5 F 13.5
8:00 221.6 F 8.0
10:00 236.3 F 13.5
12:00 238.2 F 11.1
14:00 237.4 F 9.8
16:00 251.3 F 16.6
18:00 263.1 F 23.5
20:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.6 F 29.7
2:00 271.5 F 29.0
3:00 272.5 F 29.9
4:00 272.1 F 29.4

Min 271.2 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 216.4 F 0.0
10:00 228.8 F 0.0
15:00 230.9 F 0.0
20:00 228.0 F 0.0
25:00 225.8 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:11
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

06/12/2023 12:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____