
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001657
11:05 11/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:25
STERILIZING 3:28
136.0 C 220kPa 3:28
136.4 C 226kPa 3:58
136.3 C 222kPa 4:28
136.3 C 223kPa 4:58
136.3 C 223kPa 5:28
136.3 C 224kPa 5:58
136.3 C 223kPa 6:28
136.2 C 223kPa 6:58
136.2 C 223kPa 6:58
VENTING 6:59
AIR DRYING 7:54
CYCLE COMPLETE 11:00

Digital Signature #
DFAC6E4B218716F0
