
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005250
9:13 20/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:13
STERILIZING 4:11
136.0 C 220kPa 4:11
136.3 C 223kPa 4:41
136.2 C 221kPa 5:11
136.4 C 223kPa 5:41
136.3 C 221kPa 6:11
136.4 C 222kPa 6:41
136.4 C 221kPa 7:11
136.4 C 224kPa 7:41
136.4 C 224kPa 7:41
VENTING 7:41
AIR DRYING 8:35
CYCLE COMPLETE 11:57

Digital Signature #
54C11A0AA7E648EC
