
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001157
10:16 08/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:53
STERILIZING 4:03
136.0 C 221kPa 4:03
136.0 C 222kPa 4:33
136.4 C 223kPa 5:03
136.4 C 226kPa 5:33
136.3 C 224kPa 6:03
136.4 C 226kPa 6:33
136.3 C 224kPa 7:03
136.3 C 224kPa 7:33
136.3 C 224kPa 7:33
VENTING 7:33
AIR DRYING 8:23
CYCLE COMPLETE 19:15

Digital Signature #
A96B6EAC9A5EC2E4
