

-----  
Midmark M11 - v1.0.0  
Total Cycles: 584

Name: MIDMARK  
Sterilizer ID: V2178198  
12/29/2020 09:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.0 F 0.0  
2:00 167.1 F 1.4  
4:00 168.8 F 5.2  
6:00 192.7 F 10.0  
8:00 210.4 F 4.4  
10:00 225.4 F 8.7  
12:00 236.4 F 11.0  
14:00 240.2 F 11.3  
16:00 238.6 F 9.5  
18:00 252.7 F 16.5  
20:00 264.7 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.5 F 28.5  
2:00 272.1 F 28.5  
3:00 272.3 F 28.3  
4:00 272.4 F 28.7

Min 271.6 F 28.0  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 213.1 F 0.0  
10:00 230.0 F 0.0  
15:00 235.7 F 0.0  
20:00 226.7 F 0.0  
25:00 222.8 F 0.0  
30:00 219.9 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 21:50  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:54  
CYCLE COMPLETE

12/29/2020 10:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_