

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1516

Name: RITTER 1  
Sterilizer ID: V2178198  
01/12/2022 01:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 216.6 F 0.0  
2:00 210.9 F 4.9  
4:00 220.9 F 10.8  
6:00 227.7 F 6.6  
8:00 238.0 F 12.4  
10:00 238.4 F 10.3  
12:00 237.4 F 8.9  
14:00 251.5 F 16.1  
16:00 263.6 F 23.2  
18:00 271.5 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.7 F 29.0  
2:00 271.6 F 28.3  
3:00 272.9 F 29.2  
4:00 272.1 F 28.5  
  
Min 271.5 F 28.2  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 217.9 F 0.0  
5:00 233.5 F 0.0  
10:00 239.4 F 0.0  
15:00 231.1 F 0.0  
20:00 228.3 F 0.0  
25:00 228.1 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 2:26  
HEATING: 18:02  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:40  
CYCLE COMPLETE

01/12/2022 02:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_