
Midmark M11 - v1.0.5
Total Cycles: 2988

Name: RITTER2
Sterilizer ID: V2313228
10/10/2023 12:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.1 F 0.0
2:00 154.0 F 2.6
4:00 170.8 F 5.9
6:00 191.6 F 10.2
8:00 213.3 F 5.8
10:00 220.3 F 8.4
12:00 233.2 F 13.3
14:00 235.9 F 10.3
16:00 235.6 F 8.6
18:00 249.0 F 15.2
20:00 260.4 F 21.5
22:00 270.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.3 F 29.7
2:00 271.5 F 28.9
3:00 272.1 F 29.7
4:00 271.4 F 29.0

Min 271.2 F 28.8
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 236.7 F 0.0
10:00 214.6 F 0.0
15:00 183.0 F 0.0
20:00 178.0 F 0.0
25:00 176.4 F 0.0
30:00 183.3 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:39
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

10/10/2023 01:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____