
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000182
10:43 06/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:32
PRESSURIZING 0:45
STERILIZING 1:54
136.0 C 220kPa 1:54
136.3 C 224kPa 2:24
136.2 C 221kPa 2:54
136.3 C 224kPa 3:24
136.3 C 221kPa 3:54
136.2 C 223kPa 4:24
136.4 C 224kPa 4:54
136.4 C 224kPa 5:24
136.4 C 224kPa 5:24
VENTING 5:24
AIR DRYING 6:19
CYCLE COMPLETE 6:25

Digital Signature #
051AC44FC0FD7D54
