
Midmark M11 - v1.0.0
Total Cycles: 2337

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2022 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.1 F 0.0
2:00 204.9 F 4.3
4:00 201.7 F 9.3
6:00 224.8 F 7.3
8:00 232.5 F 10.0
10:00 236.6 F 8.1
12:00 246.2 F 13.5
14:00 244.6 F 12.4
16:00 257.6 F 19.3
18:00 269.1 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 29.0
2:00 271.9 F 28.4
3:00 271.4 F 28.5
4:00 272.5 F 29.0

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 240.3 F 0.0
10:00 232.1 F 0.0
15:00 229.2 F 0.0
20:00 227.9 F 0.0
25:00 229.7 F 0.0
30:00 227.0 F 0.0

Drying Complete

FILLING: 2:08
HEATING: 19:06
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

09/19/2022 01:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____