

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3807

Name: RITTER 1  
Sterilizer ID: V2178198  
09/25/2023 10:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
2:00 182.7 F 0.1  
4:00 190.5 F 2.5  
6:00 212.2 F 6.7  
8:00 225.0 F 11.6  
10:00 228.7 F 7.4  
12:00 241.7 F 12.7  
14:00 238.4 F 10.2  
16:00 236.3 F 8.3  
18:00 249.9 F 14.9  
20:00 261.2 F 21.3  
22:00 271.5 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 271.7 F 28.5  
2:00 271.5 F 28.4  
3:00 271.7 F 28.4  
4:00 272.3 F 28.6

Min 271.5 F 28.4  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.3 F 0.0  
5:00 224.3 F 0.0  
10:00 225.2 F 0.0  
15:00 226.3 F 0.0  
20:00 217.9 F 0.0  
25:00 217.1 F 0.0  
30:00 239.2 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:13  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:01:37  
CYCLE COMPLETE

09/25/2023 11:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_