

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 001471  
11:30 10/04/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:42  
PRESSURIZING 6:42  
STERILIZING 8:05  
136.0 C 220kPa 8:05  
136.4 C 222kPa 8:35  
136.2 C 221kPa 9:05  
136.4 C 222kPa 9:35  
136.4 C 222kPa 10:05  
136.3 C 222kPa 10:35  
136.5 C 220kPa 11:05  
136.4 C 219kPa 11:35  
VENTING 11:35  
AIR DRYING 12:41  
CYCLE COMPLETE 40:29

-----  
Digital Signature #  
11AFDF857E4C0A9F  
-----