

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4980

Name: RITTER 1  
Sterilizer ID: V2178198  
10/09/2024 11:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 95.9 F 0.0  
2:00 96.2 F 0.0  
4:00 115.6 F 0.1  
6:00 143.5 F 0.7  
8:00 168.2 F 2.0  
10:00 189.9 F 4.7  
12:00 207.3 F 8.3  
14:00 217.7 F 12.0  
16:00 219.3 F 5.5  
18:00 229.8 F 9.8  
20:00 237.2 F 10.1  
22:00 239.6 F 11.5  
24:00 237.6 F 9.4  
26:00 250.6 F 15.6  
28:00 261.4 F 21.7  
30:00 271.4 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.2 F 29.0  
2:00 271.8 F 28.4  
3:00 271.2 F 28.4  
4:00 272.1 F 29.0

Min 271.0 F 28.1  
Max 272.3 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.2 F 0.0  
5:00 204.0 F 0.0  
10:00 184.8 F 0.0  
15:00 190.2 F 0.0  
20:00 201.3 F 0.0  
25:00 216.6 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:42  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:09:15  
CYCLE COMPLETE

10/09/2024 12:14 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_