
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005373
14:36 18/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:12
STERILIZING 3:14
136.4 C 222kPa 3:14
136.2 C 224kPa 3:44
136.3 C 224kPa 4:14
136.2 C 221kPa 4:44
136.2 C 224kPa 5:14
136.4 C 225kPa 5:44
136.4 C 225kPa 6:14
136.3 C 223kPa 6:44
136.3 C 223kPa 6:44
VENTING 6:44
AIR DRYING 7:35
CYCLE COMPLETE 7:40

Digital Signature #
C5FA0491FB733C
