
Midmark M11 - v1.0.5
Total Cycles: 2727

Name: RITTER2
Sterilizer ID: V2313228
08/03/2023 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 70.5 F 0.0
4:00 79.1 F 0.0
6:00 102.4 F 0.0
8:00 143.2 F 0.0
10:00 173.0 F 0.0
12:00 198.0 F 0.2
14:00 213.5 F 3.5
16:00 222.1 F 6.5
18:00 231.6 F 10.1
20:00 237.7 F 10.3
22:00 240.1 F 11.4
24:00 237.2 F 9.5
26:00 238.5 F 9.5
28:00 249.4 F 15.1
30:00 261.1 F 21.8
32:00 271.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.4 F 29.5
2:00 271.9 F 28.6
3:00 272.5 F 29.5
4:00 272.1 F 28.8

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 210.3 F 0.0
10:00 217.3 F 0.0
15:00 219.9 F 0.0
20:00 225.6 F 0.0
25:00 249.3 F 0.0
30:00 237.3 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:29
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:10:27
CYCLE COMPLETE

08/03/2023 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____