
Midmark M11 - v1.0.0
Total Cycles: 3505

Name: RITTER 1
Sterilizer ID: V2178198
07/10/2023 10:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.5 F 0.0
2:00 200.9 F 2.4
4:00 203.0 F 6.9
6:00 223.8 F 12.4
8:00 230.1 F 8.0
10:00 241.8 F 13.5
12:00 240.5 F 11.2
14:00 238.3 F 9.7
16:00 252.4 F 16.5
18:00 264.0 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.6 F 28.9
2:00 271.7 F 28.5
3:00 272.1 F 28.6
4:00 272.2 F 28.9

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 221.1 F 0.0
10:00 234.1 F 0.0
15:00 235.3 F 0.0
20:00 228.1 F 0.0
25:00 225.6 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

07/10/2023 11:34 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____